


Blood Mimicking Fluid

Model 046



The Model 046 Blood Mimicking Fluid is intended for use in any flow phantom and with any type of pumping mechanism. It was formulated to simulate the acoustic and physical characteristics of blood, thus providing a stable and reliable fluid for Doppler studies and system evaluations.

The fluid is non-hazardous and is formulated to meet the requirements for recommended blood-mimicking fluid as described in the IEC 1685 draft specifications. The fluid is fully degassed prior to packaging to minimize noise from air bubbles. In addition the scatters are neutrally buoyant thus preventing clumping and settling of the particles. Each batch of fluid is tested for speed of sound, attenuation, density and viscosity traceable to NIST.



2428 Alameda Avenue Suite 316 • Norfolk, Virginia 23513 • USA
Tel: 800.617.1177 • 757.855.2765 • Fax: 757.857.0523

WWW.CIRSINC.COM

CIRS

Tissue Simulation & Phantom Technology

MODEL 046 INCLUDES

| PROPERTY | HUMAN BLOOD (37°C) | MODEL 046 BLOOD FLUID |
|--|-----------------------|-----------------------|
| Viscosity (mPa's) | 3 | 4 ± 0.5 |
| Velocity (m s ⁻¹) | 1583 | 1570 ± 30 |
| Attenuation (dB cm ⁻¹ MHz) | 0.15 | < 0.1 |
| Backscatter (f ⁴ m ⁻¹ sr ⁻¹) | 4 X 10 ⁻³¹ | Not Measured |
| Fluid Properties | non Newtonian | Newtonian |

| QTY | COMPONENT DESCRIPTION |
|-----|-------------------------------|
| 1 | Blood Mimicking Fluid (1.8 L) |
| 1 | User Guide |
| - | 6 Month Warranty |

SAFETY INFORMATION: CIRS has evaluated this fluid and does not believe it poses a hazard at the quantities and concentrations provided. Although this product has been evaluated and found to be safe if used as intended, it is always best to avoid repeated or prolonged skin contact with materials not intended for use on the skin.

DISPOSAL INFORMATION: Solidify with vermiculite or other suitable material and then place in the trash.

STORAGE INFORMATION: Settling may occur during prolonged storage. Stir periodically to prevent clumping of particulates within this fluid.

Reference:

"Validation of a New Blood-Mimicking Fluid for Use in Doppler Flow Test Objects", K. Ramnarine, et. al., Ultrasound in Medicine & Biology, Vol. 24. No. 3, pp.454.

